

Anklizer II[®] Pediatric Walker



Figure 1



Figure 2




Figure 3



Figure 4

Please read the following instructions carefully and completely before use. Correct application is required to ensure proper function of the device.

Intended for single patient use.



WARNING: If your pain increases or persists or you experience any unusual reactions while using this product, consult your healthcare provider immediately.

APPLICATION INSTRUCTIONS

1. Unfasten straps, open liner, remove pads located inside boot and set aside. (Figure 1)

2. Wrap liner around the leg and foot ensuring the ankle pad is positioned over the top of the foot. Secure hook to the liner.

NOTE: Pads may be inserted at either side of the ankle or at heel for additional comfort.

3. Detach the figure eight strap from the parallel uprights, but do not remove completely. (Figure 2)

4. While seated, place leg in the brace and position heel at the back of the walker.

5. Position the parallel uprights to the leg per healthcare provider's recommendation. Remove plastic sleeves from the uprights to expose hook and secure to the liner.

6. Pull and attach each figure eight strap back to the parallel uprights making an "X" over the front of the ankle. (Figure 3) Slide strap back through uprights and secure back on itself.

7. Beginning with the toe strap, thread straps through links and work up the leg securing each strap.

8. Fasten closure straps until all are secured.

NOTE: Liner and strap adjustment may be necessary as swelling patterns change throughout the rehabilitation process.

NOTE: Please refer to sew-in label on device for care instructions.

NOTICE: Every effort has been made to obtain the maximum compatibility of function, strength, durability and comfort with this product, however, there is no guarantee that injury will be prevented with its use.

**NOT MADE WITH
natural rubber latex**