



Generic Tender Specifications for Height Adjustable Wellness Bathtub

Specifications for basic height adjustable bathtub

- Dual actuator electric height adjustable to enable ergonomic working methods for the caregiver and easy access from head end and either side of the bath
- Bath must be designed to work with a mobile patient lift system, chair and stretcher, to safely manage the entire bathing cycle
- Bathtub to be manufactured of double layer gel coated reinforced fiber glass with stainless steel feet
- Bathtub must include thermostatic mixing valve with preset temperature limit at 43 degrees Celsius/110 degrees Fahrenheit
- Bathtub must include a patient shower with hose and shower handle
- Bathtub must include overflow control, pop-up drainage and drainage hose, with minimum diameter of 40 mm/1^{2/3} inch and a length of a minimum of 1000 mm/39 inch
- Bathtub must be equipped with two chrome grab handles for patient comfort
- Bathtub must include a removable foot support, adjustable in three positions
- Bath to include height adjustable feet in case of uneven floor surfaces

Specification text to add for AIR SPA system

- Push button or touch pad ON/OFF controlled Air Spa system with 12 integrated air jets to stimulate blood circulation and provide a calming environment for relaxation
- The Air Spa system must include an automatic purge function to remove any remaining water from the air jets after bathing

Specification text to add for AUTOFILL system

- Push button or touch pad controlled auto fill on/off with dual preset fill levels to choose from

Specification text to add for LED light system

- Bath to include a LED Color light system that changes the water into different colors

Specification text to add for MUSIC/SPEAKER system

- Bath to include a AUX/Bluetooth music sound system with minimum 4 inbuilt speakers

Specification text to add for CLEANING system

- Integrated disinfection cleaning system

Dimension and load specifications

- | | | |
|-------------------------------------|----------------------|-------------------------------------|
| • Operating Height Range (bathtub): | Min: 650 mm/26 inch | Max: 1000 mm/39 ^{2/5} inch |
| • Length (bathtub): | Max: 2080 mm/82 inch | |
| • Length (inside bathtub) | Min: 1650 mm/65 inch | |
| • Width (bathtub): | Min: 750 mm/31 inch | Max: 780 mm/29 ^{1/2} inch |
| • Weight Limit (patient): | 150 kg/330 lbs. | |
| • Overall Weight (empty) | Max: 100 kg/220 lbs. | |
| • Min Water Volume (empty) | 180 l/48 gallons | |

