



Qty: 1

PN: 30-0088



**NORTH AMERICAN RESCUE**  
www.NARescue.com • 888.689.6277

Patent(s) Pending



**Instructions Overview**

The Junctional Emergency Treatment Tool (JETT™) includes two individually adjustable compression pads in a single tool, allowing for simultaneous occlusion of blood flow to both lower limbs (bilateral) with an easy to apply, pre-assembled, ready-to-use belt solution. The unique design of the JETT™ allows for patient transport while the device is in place and includes a lanyard and toggle solution to ensure that the Threaded T-Handle (windlass axle) does not loosen due to vibration or bumps during transport.

In cases where there are bilateral injuries, the JETT™ offers a simple, safe solution for occluding blood flow to both lower extremities. This expanded capability is built into every JETT™, meaning that with standardized training the device will be pre-positioned for rapid bilateral application whether the casualty requires a unilateral or bilateral treatment.

High extremity wounds in the thigh/groin region where tourniquets cannot be placed or are less effective require equipment such as the JETT™ to stop the bleeding and potentially save the life of a casualty that would otherwise bleed out in minutes. Penetrating trauma, blast injuries and amputation are examples of injuries that may require treatment with this device.

**Anatomy of the JETT™**



**Instructions For Use**

**Step 1: ASSESSMENT**

- 1.1 Determine that the injury is not amenable to treatment with a standard tourniquet (ie: high in the thigh and/or groin region)
- 1.2 Determine if a bilateral or unilateral application is required



**Step 2: PATIENT PREPARATION**


- 2.1 Remove any gear, items in pockets or debris that might interfere with the placement of the JETT™ or cause additional trauma.

**Step 3: DEVICE DEPLOYMENT**


- 3.1 Unroll the device in preparation for application
- 3.2 Placement of JETT™
  - 3.2.A Slide the belt with "This Side Toward Casualty" facing up under the body at the lower back
  - 3.2.B Locate the Superior Iliac Crest and Pubic Bone, which are connected by the Inguinal Ligament
  - 3.2.C Lifting the hips, position the belt under the buttocks so that the buttocks pad is centered behind the casualty and the pressure pads are immediately below (inferior to) the Inguinal Ligament




**Anatomical Landmarks for Pad Placement**



**Inguinal Ligament Identification**




# OTHER SIDE TOWARD CASUALTY



### 3.3 Initial Placement of the Pressure pads

- 3.3.A Palpate to determine the location of the Inguinal Ligament
- 3.3.B Angle the pressure pads parallel to the Inguinal Ligament (approximately 30°) pointing midline toward the feet making sure that both pads are correctly positioned
- 3.3.C Ensure that casualty's genitals are clear of the area where the pads will be positioned



3.4 Buckle the belt, then firmly tighten by pulling the loop handle while stabilizing the belt with your other hand on top of the belt on the casualty (remove ALL slack)



3.6 With one hand on the base plate, grip the threaded T-handle and rotate it clockwise to tighten



3.5 Reassess the pad placement to ensure that they are still in the correct location below the inguinal ligament

3.7 Increase pressure until bright red bleeding stops (hemostasis is achieved) and if the extremity is intact, check to confirm that distal pulse has been occluded



3.8 Insert the toggle into the opening on the threaded T-handle (pulling the handle in clockwise rotation) and cinch it tight at the base plate in order to secure the device



3.9 Repeat as necessary with the other threaded T-handle for bilateral injuries

3.10 Note time of application on Writeable Area



### Step 4: Monitoring JETT™ and Patient Reassessment

- 4.1 Assess frequently (every 5 minutes) to ensure that bleeding is still controlled (or based on your medical protocols)
- 4.2 Assess after any movement to ensure that pressure pads are in correct location and that pressure was not reduced during movement (hemostasis is maintained)
- 4.3 Device should only be removed under physician supervision (or based on your medical protocols)
- 4.4 Device application should not exceed 4 hours

#### Warning:

- Do not leave applied for longer than 4 hours
- Inguinal use only

#### Caution:

- For use by trained personnel
- Tighten until hemostasis is achieved
- Reassess casualty condition after movement or every 5 minutes (or based on your medical protocols)
- Caution: Federal law restricts this device to sale by or on the order of a physician

#### Limitation of Liability:

- North American Rescue, its employees, agents, contractors, supplier, and distributors shall assume no liability for injury or damages arising from the application and use of the Junctional Emergency Treatment Tool (JETT™). The user assumes all risk of liability.

