

LIFETIME LIMITED WARRANTY

If you have a question about your MOBB Healthcare device or this warranty, please contact an authorized MOBB Healthcare dealer.

MOBB Healthcare warrants to the original consumer purchaser for as long as they own their MOBB product that it will be free from defects in material and workmanship under normal use. This warranty is only applicable to products used in residential applications within the United States and Canada and is not transferable to subsequent owners. Should defects in material or workmanship exist during the applicable warranty period in any MOBB Healthcare Product, MOBB Healthcare will elect at its discretion to either repair or replace the defective material. This warranty covers only parts and materials of the products supplied by MOBB Healthcare. Not covered under this warranty are wear and tear items or abuse of the product in a manner that was not intended for the product to be used. This warranty shall become void if the product is in any way modified, improperly adjusted or abused.



Aluminum Wheelchair **300 lbs Weight Capacity - 18" Seat Width**



Shown in white for detail. Actual chair frame is black

INSTRUCTIONS INSIDE

SAFETY FIRST

Ensure the armrest is securely locked BEFORE using the wheelchair.

Do not stand on the frame of the wheelchair. Always use the hand rims for self propulsion. Do not grip the rear wheels themselves.

The wheelchair is suitable for indoors and out, but avoid prolonged bumpy terrain. Use proper caregiver techniques when negotiating obstacles. Consult with your MOBB Healthcare dealer for proper usage.

The wheel locks are used while the occupant is sitting and stationary, or getting into or out of the wheel chair. The wheel locks are not meant to arrest the chair while moving. Doing so could cause injury and/or wheelchair damage.

WHEELCHAIR SET UP & USAGE

1. Unfolding

- Locate the two rails on the seat rest. This is where the seat fabric is attached to. Push down on the seat rails and the chair will widen.

2. Fold

- Depress locking tabs on the back of the backrest "canes" to allow backrest to fold.
- Turn the right and left foot plates upward to the vertical position.
- Pull up on the middle of section of the cushion to fold.

3. Attach footrest

- Find the small vertical post at the top of the foot rest frame. Insert it into the corresponding hole on the frame of the wheelchair. Rotate the leg rest frame to align with the wheelchair frame and you should hear an audible click as it engages.

4. Adjustment of Foot Plate.

- Remove the screw from the foot rest frame. Slide footplate up and down to desired height. Re-install screw and nut.

WHEELCHAIR SET UP & USAGE

5. Anti-tippers

- Remove the rubber tube end where the anti-tippers will be placed.

Insert the anti-tipper housing onto the wheelchair frame and insert screw through housing and wheelchair frame. Do not over tighten.

- To adjust anti-tipper height, pull back on red tabs and adjust anti-tippers to desired height. Consult with your health care provider or wheelchair retailer for the proper height for you.



Shown in white for detail. Actual wheelchair is black.